

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 29D0931553	(X3) Date Survey Completed 05/12/2021
Name of Provider or Supplier Skin Cancer & Dermatology Institute-Gs Richards Bl	Street Address, City, State 3950 Gs Richards Blvd, Carson City, NV	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	This Statement of Deficiencies was created as a result of an on-site CLIA recertification survey conducted at your facility on 5/12/2021. The findings and conclusions of any investigation by the Division of Public and Behavioral Health shall not be construed as prohibiting any criminal or civil investigations, actions or other claims for relief that may be available to any party under applicable federal, state, or local laws.
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview with the Mohs technician, expired solutions were found in the flammable cabinet during a tour of the Mohs laboratory. Findings include: 1. An open gallon bottle of EverClear xylene substitute which expired in 2019 was available for use. 2. Four gallon-sized bottles of dehydrant, one of which was opened and in use, expired on 4/18/2021. The Mohs technician indicated during the on-site inspection on 5/12/2021 at approximately 11:30 AM that he was unaware that the bottles had already expired since he received the bottles from the supplier on 1/14/2021. The laboratory performs approximately 780 histopathology tests annually.</p>