

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 29D0931553	(X3) Date Survey Completed 07/26/2023
Name of Provider or Supplier Skin Cancer & Dermatology Institute-Gs Richards Bl	Street Address, City, State 3950 Gs Richards Blvd, Carson City, NV	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	This Statement of Deficiencies was generated as a result of the onsite CLIA recertification survey conducted at your facility on July 26, 2023. The findings and conclusions of any investigation by the Division of Public and Behavioral Health shall not be construed as prohibiting any criminal or civil investigations, actions or other claims for relief that may be available to any party under applicable federal, state or local laws.
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the laboratory's documentation of accuracy verification for provider performed microscopy (PPM) testing, and an interview with the practice manager, the laboratory failed to verify the accuracy of PPM tests for each provider least twice a year. Findings include: 1. A review of the laboratory's documentation of twice a year accuracy verification for PPM testing from January 2021 through July 2023 revealed there was only one accuracy check documented in 2022 for one of four providers. This provider is identified as number 4 on the CMS-209. 2. These findings were confirmed by the practice manager in an interview on July 26, 2023. The laboratory performs approximately 25 microbiology tests per year.</p>