

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 29D0983069	(X3) Date Survey Completed 12/05/2018
Name of Provider or Supplier Lakes Dermatology	Street Address, City, State 8861 W Sahara Ave Ste 290, Las Vegas, NV	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	<p>This Statement of Deficiencies was created as a result of an on-site CLIA recertification survey conducted at your facility on December 5, 2018. The findings and conclusions of any investigation by the Division of Public and Behavioral Health shall not be construed as prohibiting any criminal or civil investigations, actions or other claims for relief that may be available to any party under applicable federal, state, or local laws.</p>
D5203	<p>SPECIMEN IDENTIFICATION AND INTEGRITY CFR(s): 493.1232</p> <p>The laboratory must establish and follow written policies and procedures that ensure positive identification and optimum integrity of a patient's specimen from the time of collection or receipt of the specimen through completion of testing and reporting of results.</p> <p>This STANDARD is not met as evidenced by: Based on a random audit of patient histopathology records and slides from 1/05/17 through 11/08/18 and an interview with the facility nurse and the facility histotechnologist, the laboratory failed to ensure the positive identification of the patient's specimen from the time of collection through the completion of testing and reporting of results. Findings include: A random audit of six patients from 1/05/17 through 11/08/18 revealed one of the six patients who had a Mohs procedure performed on 7/15/18 had the source of the specimen to be the "right temple" on the requisition, the Mohs map and the final report. The four slides that were generated to be read by the provider, had "left temple" written on all four of the slides. This was confirmed by the facility nurse and the facility histotechnologist on 12/05/18 at approximately 3:30 pm. The laboratory performs approximately 80 patient histopathology tests annually.</p>