

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 29D1011156	(X3) Date Survey Completed 12/11/2018
Name of Provider or Supplier Las Vegas Skin & Cancer Clinics-Summerlin	Street Address, City, State 653 N Town Center Dr Ste 214, Las Vegas, NV	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	<p>This Statement of Deficiencies was created as a result of an on-site CLIA recertification survey conducted at your facility on December 11, 2018. The findings and conclusions of any investigation by the Division of Public and Behavioral Health shall not be construed as prohibiting any criminal or civil investigations, actions or other claims for relief that may be available to any party under applicable federal, state, or local laws.</p>
D5785	<p>CORRECTIVE ACTIONS CFR(s): 493.1282(b)(3)</p> <p>(b) The laboratory must document all corrective actions taken, including actions taken when any of the following occur: (b)(3) The criteria for proper storage of reagents and specimens, as specified under 493.1252(b), are not met.</p> <p>This STANDARD is not met as evidenced by: Based on a review of temperature recording logs from January 1, 2017 through December 11, 2018 and an interview with the director of compliance, the laboratory failed to document daily temperatures to monitor the proper storage of reagents and specimens. Findings include: 1. The laboratory failed to document daily temperatures for the cryostat machine of minus 20 to minus 30 degrees Centigrade, the laboratory refrigerator of 2 to 8 degrees Centigrade and the laboratory room temperature of 22 to 25 degrees Centigrade. 2. There were no temperature recordings for 7 of 19 days for the cryostat machine, the laboratory refrigerator and the laboratory room temperature for the month of September 2018. 3. There were no temperature recordings for 13 of 23 days for the cryostat machine, the laboratory refrigerator and the laboratory room temperature for the month of October 2018. 4. There were no temperature recordings for 18 of 22 days for the cryostat machine, the laboratory refrigerator and the laboratory room temperature for the month of November 2018. This was confirmed by the director of compliance on December 11, 2018 at approximately 2:30 PM. The laboratory performs approximately 475 histopathology patient tests annually.</p>