

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 29D2076312	(X3) Date Survey Completed 06/14/2019
Name of Provider or Supplier Innovative Pain Care Center-Rancho	Street Address, City, State 501 S Rancho Dr Ste G-44, Las Vegas, NV	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	<p>This Statement of Deficiencies was created as a result of an on-site CLIA recertification survey conducted at your facility on June 14, 2019. The findings and conclusions of any investigation by the Division of Public and Behavioral Health shall not be construed as prohibiting any criminal or civil investigations, actions or other claims for relief that may be available to any party under applicable federal, state, or local laws.</p>
D5203	<p>SPECIMEN IDENTIFICATION AND INTEGRITY CFR(s): 493.1232</p> <p>The laboratory must establish and follow written policies and procedures that ensure positive identification and optimum integrity of a patient's specimen from the time of collection or receipt of the specimen through completion of testing and reporting of results.</p> <p>This STANDARD is not met as evidenced by: Based on a random audit of patient laboratory testing in the areas of routine chemistry and toxicology between 7/20/17 and 5/18/19 and an interview with the laboratory manager, the laboratory failed to follow the director approved policy to ensure positive identification of a patient's specimen from the time of collection through the reporting of the test results. Findings include: 1. A random audit of patient laboratory testing in the areas of routine chemistry and toxicology between 7/20/17 and 5/18/19 found that the laboratory failed to ensure the positive identification of one of ten patient's specimens from the time of collection through the reporting of patient results. 2. The random audit found a specimen collected on 10/25/18 with a unique identification of the patient birth date of 5/23/82 had routine chemistry and toxicology testing performed on 11/03/18 with a unique identification of the patient birth date as 5/23/83 on the patient final report. This was confirmed by the laboratory manager on June 14, 2019 at approximately 3:30 PM. The laboratory performs approximately 225,000 routine chemistry and toxicology patient tests annually.</p>