

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  29D2129860	<b>(X3) Date Survey Completed</b>  01/30/2019
<b>Name of Provider or Supplier</b>  Mohave Dermatology-Pahrump	<b>Street Address, City, State</b>  1420 E Calvada Blvd Ste 100, Pahrump, NV	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	<p>This Statement of Deficiencies was created as a result of an on-site CLIA recertification survey conducted at your facility on January 30, 2019. The findings and conclusions of any investigation by the Division of Public and Behavioral Health shall not be construed as prohibiting any criminal or civil investigations, actions or other claims for relief that may be available to any party under applicable federal, state, or local laws.</p>
<b>D5217</b>	<p><b>EVALUATION OF PROFICIENCY TESTING PERFORMANCE</b> CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the twice a year verification of accuracy documentation for dermatologists performing Mohs histopathology testing and an interview with the office manager, the laboratory failed to perform and document the accuracy of dermatologists performing Mohs procedures at least twice per testing year. Findings include: 1. There was no performance and documentation of twice a year verification of accuracy for testing person number one from the Centers for Medicare &amp; Medicaid Services (CMS) 209 form for testing year 2017. 2. There was no performance and documentation of twice a year verification of accuracy for testing person number two from the (CMS) 209 form for testing year 2018. This was confirmed by the office manager on January 30, 2019 at approximately 10:00 AM. The laboratory performs approximately 96 histopathology tests annually.</p>