

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 29D2130145	(X3) Date Survey Completed 02/25/2021
Name of Provider or Supplier Velazquez Pain Relief Center - Flamingo	Street Address, City, State 8845 W Flamingo Rd Ste 100, Las Vegas, NV	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	This Statement of Deficiencies was created as a result of an on-site CLIA recertification survey conducted at your facility on February 25, 2021. The findings and conclusions of any investigation by the Division of Public and Behavioral Health shall not be construed as prohibiting any criminal or civil investigations, actions or other claims for relief that may be available to any party under applicable federal, state, or local laws.
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the American Proficiency Institute (API) proficiency testing reports for testing years 2019 and 2020, a review of proficiency testing corrective action reports for testing years 2019 and 2020 and an interview with the laboratory director, the laboratory failed to verify the accuracy at least twice annually of all non-regulated analytes performed by the laboratory. Findings include: 1. The laboratory failed to verify the accuracy at least twice during the testing year 2020 of the adulterant specific gravity. 2. The API results for specific gravity revealed a passing score of 100 % for the first event 2020 and a failing score of 67% for the second event 2020. 3. The laboratory failed to perform a second accuracy assessment for specific gravity for testing year 2020 as outlined in the director approved policy and procedure manual for failed proficiency testing. This was confirmed by the laboratory director on February 25, 2021 at approximately 10:30 am. The laboratory performs approximately 30,240 chemistry tests annually.</p>