

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 29D2140510	<b>(X3) Date Survey Completed</b> 03/29/2022
<b>Name of Provider or Supplier</b> Csl Plasma, Inc	<b>Street Address, City, State</b> 3220 S Nellis Blvd, Las Vegas, NV	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	This Statement of Deficiencies was created as a result of an on-site CLIA recertification survey conducted at your facility on March 29, 2022. The findings and conclusions of any investigation by the Division of Public and Behavioral Health shall not be construed as prohibiting any criminal or civil investigations, actions or other claims for relief that may be available to any party under applicable federal, state, or local laws.
<b>D6053</b>	<p><b>TECHNICAL CONSULTANT RESPONSIBILITIES</b> CFR(s): 493.1413(b)(9)</p> <p>The technical consultant is responsible for evaluating and documenting the performance of individuals responsible for moderate complexity testing at least semiannually during the first year the individual tests patient specimens.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the laboratory testing personnel records for training and competency assessment, a review of the completed CMS-209 form, and an interview with the Center Manager and the Assistant Manager of Quality, the technical consultant failed to ensure that semi-annual training and competency assessment was completed during the first year of employment for two of 27 testing personnel. Findings include: 1. There was no documentation of semi-annual training and competency assessment for testing personnel number 16 and testing personnel number 18 listed on the CMS-209 form during the first year of employment in the laboratory. 2. The findings were confirmed during an interview conducted with the Center Manager and the Assistant Manager of Quality on March 29, 2022 at approximately 9:30 AM. The laboratory performs approximately 77,425 chemistry tests annually.</p>