

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  29D2166555	<b>(X3) Date Survey Completed</b>  06/16/2026
<b>Name of Provider or Supplier</b>  Vivida Dermatology(Henderson)	<b>Street Address, City, State</b>  1736 W Horizon Ridge Pkwy, Henderson, NV	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	This statement of deficiencies was created as a result of an on site CLIA recertification survey conducted at your facility on June 16, 2026. The findings and conclusions of any investigation by the Division of Health Care Purchasing and Compliance shall not be construed as prohibiting any criminal or civil investigations, actions, or other claims for relief that may be available to any party under applicable federal, state or local laws.
<b>D6120</b>	<p><b>TECHNICAL SUPERVISOR RESPONSIBILITIES</b> CFR(s): 493.1451(b)(7)(8)</p> <p>(b)(7) Identifying training needs and assuring that each individual performing tests receives regular in-service training and education appropriate for the type and complexity of the laboratory services performed; (b)(8) Evaluating the competency of all testing personnel and assuring that the staff maintain their competency to perform test procedures and report test results promptly, accurately and proficiently.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the staff roster as documented on the CMS-209 form, the personnel competency assessments, and an interview with the laboratory manager, the technical supervisor failed to ensure that competency assessments for one of six testing personnel performing potassium hydroxide (KOH) were completed for two of two years. Findings include: 1. A review of the staff roster as documented on the CMS-209 form and the personnel competency assessments found that annual competency assessments from December 2024 and December 2025 for testing personnel (TP) number 6 on the CMS-102 had been signed, but the boxes indicating what items had been assessed and deemed competent had not been completed. 2. An interview with the laboratory manager on June 6, 2026, at approximately 2:30 pm confirmed these findings. According to the CMS-116 submitted at the time of survey, the laboratory performs 65 KOH tests annually.</p>