

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  29D2232333	<b>(X3) Date Survey Completed</b>  04/15/2026
<b>Name of Provider or Supplier</b>  Rapid Lab	<b>Street Address, City, State</b>  2465 Reynolds Ave Ste 101, N Las Vegas, NV	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	This Statement of Deficiencies was created as a result of an on-site CLIA recertification survey conducted at your facility on April 15, 2026. The findings and conclusions of any investigation by the Division of Healthcare Purchasing and Compliance shall not be construed as prohibiting any criminal or civil investigations, actions or other claims for relief that may be available to any party under applicable federal, state, or local laws.
<b>D5805</b>	<p>TEST REPORT CFR(s): 493.1291(c)</p> <p>(c) The test report must indicate the following: (c)(1) For positive patient identification, either the patient's name and identification number, or a unique patient identifier and identification number. (c)(2) The name and address of the laboratory location where the test was performed. (c)(3) The test report date. (c)(4) The test performed. (c)(5) Specimen source, when appropriate. (c)(6) The test result and, if applicable, the units of measurement or interpretation, or both. (c)(7) Any information regarding the condition and disposition of specimens that do not meet the laboratory's criteria for acceptability.</p> <p>This STANDARD is not met as evidenced by: Based on a random audit of eight patient records from July 3, 2024 through February 13, 2026, and interviews with the laboratory manager and laboratory director, the laboratory failed to ensure that the Estimated Glomerular Filtration Rate (eGFR) was calculated using the race-neutral equation after December 28, 2024 for three of the eight patient reports reviewed. Findings include: 1. A random audit of eight patient records revealed that three eGFR results dated March 18, 2025; June 4, 2025; and February 13, 2026 were not calculated using the race-neutral equation, as required after December 28, 2024. 2. This finding was confirmed during interviews with the</p>

laboratory manager and laboratory director on April 15, 2026 at approximately 2:45 PM. According to the submitted CMS-116, the laboratory performs 80,000 chemistry tests annually