

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 30D0979580	(X3) Date Survey Completed 12/11/2024
Name of Provider or Supplier Mid-State Health Center	Street Address, City, State 100 Robie Rd, Bristol, NH	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5439	<p>CALIBRATION AND CALIBRATION VERIFICATION CFR(s): 493.1255(b)</p> <p>Unless otherwise specified in this subpart, for each applicable test system the laboratory must do the following: Perform and document calibration verification procedure - (b)(1) Following the manufacturer's calibration verification instructions; (b)(2) Using the criteria verified or established by the laboratory under 493.1253(b)(3) -- (b)(2)(i) Including the number, type, and concentration of the materials, as well as acceptable limits for calibration verification; and (b)(2)(ii) Including at least a minimal (or zero) value, a mid-point value, and a maximum value near the upper limit of the range to verify the laboratory's reportable range of test results for the test system; and (b)(3) At least once every 6 months and whenever any of the following occur: (b)(3)(i) A complete change of reagents for a procedure is introduced, unless the laboratory can demonstrate that changing reagent lot numbers does not affect the range used to report patient test results, and control values are not adversely affected by reagent lot number changes. (b)(3)(ii) There is major preventive maintenance or replacement of critical parts that may influence test performance. (b)(3)(iii) Control materials reflect an unusual trend or shift, or are outside of the laboratory's acceptable limits, and other means of assessing and correcting unacceptable control values fail to identify and correct the problem. (b)(3)(iv) The laboratory's established schedule for verifying the reportable range for patient test results requires more frequent calibration verification.</p> <p>This STANDARD is not met as evidenced by: Based on record review and staff interview, the laboratory (lab) failed to perform calibration verification every six months for 2 routine chemistry tests in 2023 and 2024. Findings include: 1. Review on 12/11/2024 of calibration records for prostate specific antigen (PSA) and ferritin revealed each test is calibrated using 2 levels. The</p>

lab had no record of calibration verification every six months for PSA and ferritin. 2. Interview on 12/11/2024 at 11:00 a.m. with the Technical Consultant revealed the lab had not performed calibration verification for PSA and ferritin in 2023 and 2024.