

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 30D1057306	(X3) Date Survey Completed 03/11/2022
Name of Provider or Supplier D'Adamo Institute For The Advancement Of Natural	Street Address, City, State 44-54 Bridge Street, Portsmouth, NH	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5211	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on record review and staff interview the laboratory (lab) failed to evaluate immunohematology proficiency testing (PT) results for 3 of 3 PT events in 2021. Findings include: 1. Review on 3/11/2022 of the lab's PT records from 2020 and 2021 revealed American Proficiency Institute (API) evaluation reports, for 3 of 3 PT events, had not been dated and signed as having been reviewed. Further review of API evaluation reports revealed the lab obtained an unsatisfactory score (80%) for API event #1 of 2021 and the lab did not investigate the cause. 2. Interview on 3/11/2022 at 9:30 a.m. with the lab's owner confirmed the above finding.</p>