

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D0104198	(X3) Date Survey Completed 06/30/2022
Name of Provider or Supplier Regional Cancer Care Associates	Street Address, City, State 1 Bay Avenue, Montclair, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5439	<p>CALIBRATION AND CALIBRATION VERIFICATION CFR(s): 493.1255(b)</p> <p>Unless otherwise specified in this subpart, for each applicable test system the laboratory must do the following: Perform and document calibration verification procedure - (b)(1) Following the manufacturer's calibration verification instructions; (b)(2) Using the criteria verified or established by the laboratory under 493.1253(b)(3) -- (b)(2)(i) Including the number, type, and concentration of the materials, as well as acceptable limits for calibration verification; and (b)(2)(ii) Including at least a minimal (or zero) value, a mid-point value, and a maximum value near the upper limit of the range to verify the laboratory's reportable range of test results for the test system; and (b)(3) At least once every 6 months and whenever any of the following occur: (b)(3)(i) A complete change of reagents for a procedure is introduced, unless the laboratory can demonstrate that changing reagent lot numbers does not affect the range used to report patient test results, and control values are not adversely affected by reagent lot number changes. (b)(3)(ii) There is major preventive maintenance or replacement of critical parts that may influence test performance. (b)(3)(iii) Control materials reflect an unusual trend or shift, or are outside of the laboratory's acceptable limits, and other means of assessing and correcting unacceptable control values fail to identify and correct the problem. (b)(3)(iv) The laboratory's established schedule for verifying the reportable range for patient test results requires more frequent calibration verification.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the Sysmex Certificate of Calibrations (COC) and interview with the Testing Personnel (TP), the laboratory failed to perform calibration verification every 6 months on the Sysmex XN-430 analyzer used for Hematology testing from October 2021 to the date of survey. The findings include: 1. There was no documentation of the number, type and concentration of the materials used for performing Calibration Verification (CV). 2. The COC did not provide lot numbers of</p>

calibration material used for CV. 3. The TP confirmed on 6/30/22 at 1:20 pm that the laboratory failed to perform CV every 6 months.

D6013

LABORATORY DIRECTOR RESPONSIBILITIES

CFR(s): 493.1407(e)(3)(ii)

The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(3) Ensure that-- (e)(3)(ii) Verification procedures used are adequate to determine the accuracy, precision, and other pertinent performance characteristics of the method;

This STANDARD is not met as evidenced by:

Based on surveyor review of the Performance Specification (PS) records and interview with the Testing Personnel (TP), the Laboratory Director (LD) failed to ensure that PS procedures performed on the Sysmex XN-430 analyzer were adequate from October 2021 to the date of survey. The findings include: 1. There was no evidence that a Linearity was performed. 2. The TP confirmed on 6/30/22 at 1:35 am that PS records were not adequate.