

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D0113432	(X3) Date Survey Completed 04/15/2021
Name of Provider or Supplier Westwood Dermatology	Street Address, City, State 390 Old Hook Road, Westwood, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor observation of Histology reagents and interview with the Technical Supervisor (TS), the laboratory failed to discard expired Histopathology reagents from 3/1/21 to the date of survey. The findings include: 1. On the date of the survey the laboratory used expired reagents as follows: a. 1% Acid Alcohol Lot 077924 expiration 3/1/21. b. Scott's tap water Bluing Solution Lot J080-01 expiration 3/22/21. 2. The TS confirmed on 4/15/21 at 1:30 pm that the laboratory used expired reagents.</p>