

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D0118747	(X3) Date Survey Completed 01/25/2021
Name of Provider or Supplier Optum Medical Care Of Nj Pc	Street Address, City, State 69 North Main Street, Medford, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D1000	<p>CERTIFICATE OF WAIVER TESTS CFR(s): 493.15(c)</p> <p>Certificate of waiver tests. A laboratory may qualify for a certificate of waiver under section 353 of the PHS Act if it restricts the tests that it performs to one or more of the following tests or examinations (or additional tests added to this list as provided under paragraph (d) of this section) and no others: (1) Dipstick or Tablet Reagent Urinalysis (non-automated) for the following: (i) Bilirubin; (ii) Glucose; (iii) Hemoglobin; (iv) Ketone; (v) Leukocytes; (vi) Nitrite; (vii) pH; (viii) Protein; (ix) Specific gravity; and (x) Urobilinogen. (2) Fecal occult blood; (3) Ovulation tests-visual color comparison tests for human luteinizing hormone; (4) Urine pregnancy tests - visual color comparison tests; (5) Erythrocyte sedimentation rate-non-automated; (6) Hemoglobin-copper sulfate-non-automated; (7) Blood glucose by glucose monitoring devices cleared by the FDA specifically for home use; (8) Spun microhematocrit; and (9) Hemoglobin by single analyte instruments with self-contained or component features to perform specimen/reagent interaction, providing direct measurement and readout.</p> <p>This STANDARD is not met as evidenced by: The laboratory was found in compliance with the CLIA regulations with no deficiencies.</p>