

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 31D0119315	<b>(X3) Date Survey Completed</b> 09/11/2024
<b>Name of Provider or Supplier</b> Dermatology Center Of Washington Township Pc	<b>Street Address, City, State</b> 100 Kings Way East, Sewell, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5401</b>	<p><b>PROCEDURE MANUAL</b> CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the Procedure Manual (PM), observation of Staining Station (SS) and interview with the Testing Personnel (TP), the laboratory failed to follow the PM for "Mohs Tissue Procedure" from 1/4/24 to the date of the survey. The findings include: 1. The PM stated step i "Eosin working solutions 2 dips" but the SS had 100% alcohol 2. The TP confirmed on 9/11/24 at 1:30 pm that the laboratory did not follow the PM.</p>
<b>D5417</b>	<p><b>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT</b> CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor observation of Histology reagents and interview with the Testing Personnel (TP), the laboratory failed to discard expired Histopathology reagents from 3/31/24 to the date of survey. The findings include: 1. Avantik Tissue marking dyes were expired as follows: a) Blue marking day lot # 145362 expired 3/31/24 b) Red marking dye lot # 145595 expired 3/31/24 2. Approximately 500 patients were tested</p>

with expired reagent. 3. The TP confirmed on 9/11/24 at 1:30 pm that the laboratory used expired reagents.