

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  31D0120827	<b>(X3) Date Survey Completed</b>  11/20/2018
<b>Name of Provider or Supplier</b>  Internal Medicine Associates, Pa	<b>Street Address, City, State</b>  201 Laurel Heights Drive, Bridgeton, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5209</b>	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on lack of the Competency Assessment (CA) records, review of the procedure manual and interview with the Testing Personnel (TP), the laboratory failed to perform a CA on one out of one testing personnel in the calendar year 2017. The TP confirmed on 11/20/18 at 10:25 am that CA procedure was not followed.</p>
<b>D5421</b>	<p>ESTABLISHMENT AND VERIFICATION OF PERFORMANCE CFR(s): 493.1253(b)(1)</p> <p>Each laboratory that introduces an unmodified, FDA-cleared or approved test system must do the following before reporting patient test results: (1)(i) Demonstrate that it can obtain performance specifications comparable to those established by the manufacturer for the following performance characteristics: (1)(i)(A) Accuracy. (1)(i)(B) Precision. (1)(i)(C) Reportable range of test results for the test system. (1)(ii) Verify that the manufacturer's reference intervals (normal values) are appropriate for the laboratory's patient population.</p> <p>This STANDARD is not met as evidenced by: Based on review of the Performance Specification records and interview with the Testing Personnel (TP), the laboratory failed to verify Reference Interval (RI) for Thyroxine, Thyroid-Stimulating Hormone, Total Thyroxine, Prostate-Specific Antigen performed on the Tosoh AIA-360 analyzer from August 2017 to the date of</p>

survey. The TP confirmed on 11/20/18 at 12:00 pm that the laboratory did not perform RI verification.