

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D0123548	(X3) Date Survey Completed 10/01/2024
Name of Provider or Supplier Pediatric Affiliates, Pa	Street Address, City, State 40 Bey Lea Road, Toms River, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor observation of Bacitracin Discs (BD) and interview with the Testing Personnel (TP), the laboratory failed to discard expired (BD) from 9/30/24 to the date of survey. The findings include: 1. On the 10/1/24 the laboratory had expired BD Lot # 3062072 expired 9/30/24 2. The TP confirmed on 10/1/24 at 11:00 am that the laboratory Failed to discard expired BD.</p>