

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D0124054	(X3) Date Survey Completed 02/10/2022
Name of Provider or Supplier Tiefenbrunn & Fortin Pediatrics	Street Address, City, State 503 Cranbury Road, East Brunswick, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5415	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(c)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies, as appropriate, must be labeled to indicate the following: (1) Identity and when significant, titer, strength or concentration. (2) Storage requirements. (3) Preparation and expiration dates. (4) Other pertinent information required for proper use.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor observation of the reagents and interview with the Testing Personnel (TP), the laboratory failed to label the reagents used to perform Bacteriology testing with pertinent information required for proper use on the date of survey. The finding includes: 1. The laboratory did not document received date on Bacitracin TAXO discs lot # 0343425. 2. The TP confirmed on 2/10/22 at 10:05 am the laboratory did not document received date on Bacitracin TAXO discs lot # 0343425.</p>