

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 31D0125381	<b>(X3) Date Survey Completed</b> 11/01/2018
<b>Name of Provider or Supplier</b> Advocare North Brunswick Pediatrics	<b>Street Address, City, State</b> 1950 State Route 27, North Brunswick, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2015</b>	<p>TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b)(5)(6)</p> <p>(5) The laboratory must document the handling, preparation, processing, examination, and each step in the testing and reporting of results for all proficiency testing samples. The laboratory must maintain a copy of all records, including a copy of the proficiency testing program report forms used by the laboratory to record proficiency testing results including the attestation statement provided by the PT program, signed by the analyst and the laboratory director, documenting that proficiency testing samples were tested in the same manner as patient specimens, for a minimum of two years from the date of the proficiency testing event. (6) PT is required for only the test system, assay, or examination used as the primary method for patient testing during the PT event.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the Proficiency Testing (PT) records and interview with the Laboratory Director (LD), the laboratory failed to maintain the Attestation Statements (AS) signed by the analyst for tests performed with the College of American Pathologists (CAP) in the calendar years 2017 and 2018. The findings include: 1. The AS was not signed by the analyst for event MCS - A 2017 and MCS - A and B 2018. 2. The LD confirmed on 11/1/18 at 11:20 am that AS were not signed by the analyst.</p>
<b>D5413</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(b)</p> <p>The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, accurate and reliable test system operation, and test result reporting. The criteria must be consistent with the manufacturer's instructions, if provided. These conditions must be monitored and documented and, if</p>

applicable, include the following: (1) Water quality. (2) Temperature. (3) Humidity. (4) Protection of equipment and instruments from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports.

This STANDARD is not met as evidenced by:

Based on surveyor review of the Procedure Manual (PM), Temperature Log (TL) and interview with Laboratory Director (LD), the laboratory failed to monitor and document Room Temperature (RT) where Throat Culture tests were performed from 11/1/16 to the date of survey. The LD confirmed on 11/1/18 at 11:40 am that the laboratory did not document RT.

**D5805**

**TEST REPORT**

CFR(s): 493.1291(c)

The test report must indicate the following: (c)(1) For positive patient identification, either the patient's name and identification number, or a unique patient identifier and identification number. (c)(2) The name and address of the laboratory location where the test was performed. (c)(3) The test report date. (c)(4) The test performed. (c)(5) Specimen source, when appropriate. (c)(6) The test result and, if applicable, the units of measurement or interpretation, or both. (c)(7) Any information regarding the condition and disposition of specimens that do not meet the laboratory's criteria for acceptability.

This STANDARD is not met as evidenced by:

Based on surveyor review of the Final Report (FR) and interview with the Laboratory Director (LD), the laboratory failed to ensure that the Test Report Date (TRD) was indicated on the FR for Throat Cultures from 11/1/16 to the date of survey. The LD confirmed on 11/1/18 at 12:35 pm that the TRD was not on the FR.

**D5891**

**POSTANALYTIC SYSTEMS QUALITY ASSESSMENT**

CFR(s): 493.1299(a)

The laboratory must establish and follow written policies and procedures for an ongoing mechanism to monitor, assess and, when indicated, correct problems identified in the postanalytic systems specified in 493.1291.

This STANDARD is not met as evidenced by:

Based on surveyor review of the Procedure Manual and interview with the Laboratory Director (LD), the laboratory failed to have a procedure to verify manually entered results into Electronic Medical Records (EMR) from 11/1/16 to the date of the survey. The LD confirmed on 11/1/18 at 1:15 pm that the laboratory did not have the procedure mentioned above.

**D6018**

**LABORATORY DIRECTOR RESPONSIBILITIES**

CFR(s): 493.1407(e)(4)(iii)

The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory

director must-- (e)(4)(iii) Ensure that all proficiency testing reports received are reviewed by the appropriate staff to evaluate the laboratory's performance and to identify any problems that require corrective action;

This STANDARD is not met as evidenced by:

Based on surveyor review of Proficiency Testing (PT) records and interview with the Laboratory Director (LD), the LD failed to ensure that all PT results received were reviewed by the appropriate staff to evaluate the laboratory's performance and to identify any problems that require corrective action for Throat Culture testing performed with the College of American Pathologists (CAP) in the calendar years 2017 and 2018. The LD confirmed on 11/1/18 at 12:30 pm that the CAP PT results were not reviewed.