Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D0653051	(X3) Date Survey Completed 10/10/2019
Name of Provider or Supplier Burlington County Health Department	**Street Address, City, State** 15 Pioneer Boulevard, Westampton, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5221	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(d) All proficiency testing evaluation and verification activities must be documented. This STANDARD is not met as evidenced by: Based on surveyor review of the Proficiency Testing (PT) records and interview with the Office Manager (OM), the laboratory failed to evaluate results when they received an unacceptable score in Gram Stain tests performed with the the Collage of American Pathologists for first event 2018. The finding include: 1. The laboratory received an "Unacceptable" result Gram Stain on samples D5-05 in 1-2018. 2. There was no documented evidence that the laboratory investigated the failure. 3. The OM confirmed on 10/10/19 at 1:10 pm that the laboratory did not perform and document an evaluation of unacceptable PT results.