

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  31D0670316	<b>(X3) Date Survey Completed</b>  08/09/2018
<b>Name of Provider or Supplier</b>  Valley Pediatric Assoc Pa	<b>Street Address, City, State</b>  201 East Franklin Tpke, Ho Ho Kus, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of work records and interview with the Registered Nurse (RN), the laboratory used expired Throat Culture (TC) media when they have exceeded their expiration dates from 7/17/18 to the date of survey. The findings include: 1. TC media with lot # 8109605 Expired on 7/17/18 were used from 7/17/18 to 8/9/18. 2. Approximately 99 patients specimens were run and reported. 3. The RN confirmed on 8/9/16 at 1:00 pm that the laboratory used expired TC media.</p>