

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  31D0685177	<b>(X3) Date Survey Completed</b>  07/26/2023
<b>Name of Provider or Supplier</b>  Affiliated Dermatologists &	<b>Street Address, City, State</b>  182 South Street, Morristown, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor observation of Histopathology reagents and interview with the Testing Personnel (TP), the laboratory failed to discard expired Histopathology reagents from 6/30/23 to the date of survey. The finding include: 1. One Yellow Tissue Marking Dye Lot 123412 expired on 6/30/23. 2. No patients were tested with the expired reagent. 3. The TP #2 listed on the CMS-209 form confirmed on 7/26/23 at 10:30 am that the laboratory did not discard the expired reagent.</p>