Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 31D0685379	**(X3) Date Survey Completed** 07/24/2018
Name of Provider or Supplier Rutgers Rwjms Center For Dermatology	**Street Address, City, State** 1 World'S Fair Drive, Somerset, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D5209	PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235 As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency. This STANDARD is not met as evidenced by: Based on surveyor review of the Competency Assessment (CA) records and interview with the Testing Personnel (TP), the laboratory failed to perform a CA on one out of two TP for the calendar year 2016 and 2017. The TP #1 listed on CMS form 209 confirmed on 7/24/18 at 1:05 pm that the CA was not performed.
D5217	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1) At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part. This STANDARD is not met as evidenced by: Based on surveyor review of the Biannual Assessment (BA) records and interview with the Testing Personnel (TP), the laboratory failed to verify the accuracy of Histopathology testing twice annually in the calendar year 2017. The finding includes: 1) The laboratory performed one BA in 2017. 2) The TP#1 listed on CMS form 209 confirmed on 7/24/18 at 11:00 am the laboratory did not perform BA twice annually.