

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D0882201	(X3) Date Survey Completed 02/09/2023
Name of Provider or Supplier Summit Medical Group	Street Address, City, State 15000 Midlantic Drive, Mount Laurel, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2000	<p>ENROLLMENT AND TESTING OF SAMPLES CFR(s): 493.801</p> <p>Each laboratory must enroll in a proficiency testing (PT) program that meets the criteria in subpart I of this part and is approved by HHS. The laboratory must enroll in an approved program or programs for each of the specialties and subspecialties for which it seeks certification. The laboratory must test the samples in the same manner as patients' specimens. For laboratories subject to 42 CFR part 493 published on March 14, 1990 (55 FR 9538) prior to September 1, 1992, the rules of this subpart are effective on September 1, 1992. For all other laboratories, the rules of this subpart are effective January 1, 1994.</p> <p>This CONDITION is not met as evidenced by: Based on surveyor review of the Proficiency Testing (PT) records and interview with the Testing Personnel (TP), the laboratory failed to enroll in an approved PT program for Urine Chemistry tests for two out of three PT events in 2022. The TP confirmed on 2/9/23 at 1:00 pm the laboratory did not enrolled in PT for Urine Chemistry tests.</p>
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS</p>

may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.

This CONDITION is not met as evidenced by:
Based on surveyor review of the Proficiency Testing (PT) evaluation records, work records and interview with Testing Personnel (TP) the laboratory failed to participate in PT 3rd event 2020 Hematology/Coagulation with the American Proficiency Institute (API).

D5221

EVALUATION OF PROFICIENCY TESTING PERFORMANCE
CFR(s): 493.1236(d)

All proficiency testing evaluation and verification activities must be documented.

This STANDARD is not met as evidenced by:
Based on surveyor review of the Proficiency Testing (PT) records and interview with the Testing Personnel (TP) the laboratory failed to review and evaluate results when they received an unacceptable score in Hematology/Coagulation tests performed with the American Proficiency Institute (API), for 1st event 2021. The findings include: 1. The laboratory received an 0% grade for Hemoglobin, Glucose, Ketones, Leukocyte, Nitrite, and Protein. 2. There was no documented evidence that the laboratory ran the PT samples and compared them with API PT results. 3. The TP confirmed on 2/9/23 at 12:45 pm that the laboratory did not review and document an evaluation of unacceptable PT results.

D5469

CONTROL PROCEDURES
CFR(s): 493.1256(d)(10)(g)

Unless CMS Approves a procedure, specified in Appendix C of the State Operations Manual (CMS Pub. 7), that provides equivalent quality testing, the laboratory must-- Establish or verify the criteria for acceptability of all control materials. (i) When control materials providing quantitative results are used, statistical parameters (for example, mean and standard deviation) for each batch and lot number of control materials must be defined and available. (ii) The laboratory may use the stated value of a commercially assayed control material provided the stated value is for the methodology and instrumentation employed by the laboratory and is verified by the laboratory. (iii) Statistical parameters for unassayed control materials must be established over time by the laboratory through concurrent testing of control materials having previously determined statistical parameters. (g) The laboratory must document all control procedures performed.

This STANDARD is not met as evidenced by:
Based on surveyor review of Quality Control (QC) records and interview with the Testing Personnel (TP), the laboratory failed to verify commercially assayed QC material with each new lot and/or shipment for Urinalysis testing performed on the

Clinitek Advantus analyzer from 5/23/18 to the date of survey. The TC confirmed on 2/9/23 at 12:45 pm the assayed values of QC material were not verified before putting in use. Note: This was previously cited 10/16/19.

D5891

POSTANALYTIC SYSTEMS QUALITY ASSESSMENT
CFR(s): 493.1299(a)

The laboratory must establish and follow written policies and procedures for an ongoing mechanism to monitor, assess and, when indicated, correct problems identified in the postanalytic systems specified in 493.1291.

This STANDARD is not met as evidenced by:
Based on surveyor review of the Procedure Manual (PM) and interview with the Testing Personnel (TP), the laboratory failed to establish a procedure for verifying manually entered results from 10/16/19 to the date of survey. The TP confirmed on 2/9/23 at 1:00 pm that the laboratory did not have the procedure mentioned above.

D6029

LABORATORY DIRECTOR RESPONSIBILITIES
CFR(s): 493.1407(e)(11)

The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(11) Ensure that prior to testing patients' specimens, all personnel have the appropriate education and experience, receive the appropriate training for the type and complexity of the services offered, and have demonstrated that they can perform all testing operations reliably to provide and report accurate results.

This STANDARD is not met as evidenced by:
Based on surveyor review of the Personnel Files (PF) and interview with the Testing Personnel (TP), the Laboratory Director (LD) failed to have appropriate education documentation on all Testing Personnel (TP) performing laboratory testing on the date of survey. The findings include: 1. The laboratory did not have education records for six out of nine TP listed on the CMS form 209. 2. The TP confirmed on 2/9/23 at 12:40 pm the above records were not on file.