

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D0940452	(X3) Date Survey Completed 01/23/2019
Name of Provider or Supplier Satya P Kastuar Md	Street Address, City, State 2480 Route 27, North Brunswick, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5401	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the Procedure Manual (PM) and interview with the Laboratory Director (LD), the laboratory failed to follow the Turn Around Time (TAT) sign out procedure from 3/15/17 to the date of the survey. The finding includes: 1. The PM stated "The Specimen Accession Log (SAL) will keep track of specimen collected date and slide receipt date" but a review of the SAL revealed the slide receipt date was not recorded in the log. 2. The LD confirmed on 1/23/19 at 10:30 am that the laboratory did not follow the PM.</p>