

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D0946119	(X3) Date Survey Completed 03/13/2018
Name of Provider or Supplier Pediatric Affiliates, Pa	Street Address, City, State 1001 Rt 9n, Howell, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6046	<p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(8)</p> <p>(b) The technical consultant is responsible for-- (b)(8) Evaluating the competency of all testing personnel and assuring that the staff maintain their competency to perform test procedures and report test results promptly, accurately and proficiently.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the Personnel Files and interview with the Office Manager (OM), the Technical Consultant (TC) failed to evaluate the competency of seven out of seven Testing Personal (TP) in the calendar years 2016 and 2017. The finding includes: 1 .The OM had a high school diploma and was not qualified to perform CA 2. The OM confirmed on 2/8/18 at 1:40 pm that the TC did not perform CA.</p>