

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D0946119	(X3) Date Survey Completed 08/11/2026
Name of Provider or Supplier Pediatric Affiliates, Pa	Street Address, City, State 1001 Rt 9n, Howell, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5215	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(2) The laboratory must verify the accuracy of any analyte, specialty or subspecialty assigned a proficiency testing score that does not reflect laboratory test performance (that is, when the proficiency testing program does not obtain the agreement required for scoring as specified in subpart I of this part, or the laboratory receives a zero score for nonparticipation, or late return or results). This STANDARD is not met as evidenced by: Based on surveyor review of the Proficiency Testing (PT) records and interview with the Office Manager (OM), the laboratory failed to verify the accuracy of Bacteriology PT results from the American Proficiency Institute (API) for the 3rd event of 2025. The finding includes: 1. There was no evidence of the laboratory verifying the accuracy of Bacteriology analytes when the laboratory received a 0% Failure to participate for analytes throat culture and Urine colony count.. 2. The OM confirmed on 8/11/26 at 11:20 am the accuracy of the PT results were not verified
D6016	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i) (e)(4)(i) The proficiency testing samples are tested as required under Subpart H of this part; This STANDARD is not met as evidenced by: Based on surveyor review of the Proficiency Testing (PT) records and interview with Office Manager (OM), the Laboratory Director (LD) failed to ensure PT samples were tested for Bacteriology from the American Proficiency Institute (API) for the 3rd event of 2025. The findings include: 1) The laboratory received a 0% and a notation:

failure to participate or 3rd event of 2025 2) The OM stated "They missed the date deadline to report their findings" 3) The OM confirmed on 8/11/26 at 11:45 am that the LD did not ensure PT samples for Bacteriology were tested for the 3rd event of 2025.