

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D1006183	(X3) Date Survey Completed 08/21/2018
Name of Provider or Supplier Atlantic Dermatology Associates Llc	Street Address, City, State 1031 McBride Avenue, Woodland Park, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6091	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(4)(iii)</p> <p>The laboratory director must ensure all proficiency testing reports received are reviewed by the appropriate staff to evaluate the laboratory's performance and to identify any problems that require corrective action.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the Biannual Assessment (BA) records and interview with the Testing Personnel (TP), the Laboratory Director failed to ensure that results of Histopathology tests received from the reviewer were reviewed and evaluated for the calendar years 2016 and 2017. The TP confirmed on 8/21/18 at 10:15 am that the BA records were not reviewed and evaluated.</p>