

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D1030752	(X3) Date Survey Completed 07/12/2023
Name of Provider or Supplier Princeton Dermatology Associates	Street Address, City, State 1950 State Hwy 27, North Brunswick, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the Procedure Manual (PM), Biannual Assessment (BA) records and interview with the Testing Personnel (TP), the laboratory failed to verify the accuracy and reliability of Moh's testing twice a year from 8/10/21 until the date of survey. The finding includes: 1. The BA assessments did not have documented proof the reviewing physician was qualified to participate in the BA. 2. The TP confirmed on 7/12/23 at 1:30 pm that the laboratory did not verify the accuracy of Moh's testing twice a year.</p>
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor observation of the Flammable Cabinet and interview with the Testing Personnel (TP), the laboratory used expired 95% Reagent Alcohol for Histopathology testing from 5-31-2023 to the date of survey. The findings include: 1) StatLab 95% Reagent Alcohol lot # 122818 expired 5/31/23. 2) Approximately 35 patient were run and reported. 3) The TP confirmed on 7/12/23 at 1:30 pm that the laboratory used expired reagent.</p>