

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D1036684	(X3) Date Survey Completed 10/27/2022
Name of Provider or Supplier Atlantic Dermatology & Laser Center	Street Address, City, State 1401 New Road, Linwood, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5401	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the Procedure Manual (PM), observation of the Microscope, lack microscope maintenance records records and interview with the Office Manager (OM), the laboratory failed to follow their policy for "Equipment Quality Control - Microscope Use Protocol" in the calendar years 2020, 2021 and 2022. The finding includes: 1. The "Equipment Quality Control - Microscope Use Protocol" stated "2. Microscope gets preventative maintenance once yearly" 2. There was no documented evidence that the above mentioned procedure was performed. 3. The OM confirmed on 10/27/22 at 10:30am the laboratory did not follow the PM.</p>