

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D1040720	(X3) Date Survey Completed 10/24/2023
Name of Provider or Supplier Cosmetic & Procedural Dermatology	Street Address, City, State 215 Morris Ave, Spring Lake, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor observation of Histopathology reagents and dyes, surveyor review of the Procedure Manual (PM) and interview with the Laboratory Director (LD), the laboratory failed to dispose expired Histopathology reagents and dyes from 1/31/22 to the date of survey. The findings include: 1. The Procedure Manual (PM) states "Expired reagent chemicals are not permitted for use during frozen section processing. Any reagent that has reached the expiration date must be properly discarded. No expired chemical reagent should ever be used or stored in the Moh's laboratory beyond the expiration date." 2. One Red Tissue Marking Dye 2 ounce container Lot # 093059 expired on 1/31/22. 3. The laboratory performed approximately 1300 Histopathology tests annually. 4. The LD confirmed on 10/24/23 at 11:00 am that the laboratory failed to dispose the expired Tissue Marking Dye.</p>