

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D1058391	(X3) Date Survey Completed 04/25/2023
Name of Provider or Supplier Margaret Ravits & Associates	Street Address, City, State 721 Summit Avenue, Hackensack, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6091	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(4)(iii)</p> <p>The laboratory director must ensure all proficiency testing reports received are reviewed by the appropriate staff to evaluate the laboratory's performance and to identify any problems that require corrective action.</p> <p>This STANDARD is not met as evidenced by: Based on the lack of the Biannual Assessment (BA) records and interview with the Testing Personnel (TP), the Laboratory Director failed to ensure BA was performed to evaluate Mart-1 testing performance accurately from May 2022 to the date of survey. The TP confirmed on 4/25/23 at 11:00 am the BA was not performed to evaluate the laboratory's performance accurately.</p>