

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D1063037	(X3) Date Survey Completed 02/26/2019
Name of Provider or Supplier Nj Certified Dermatology, Pc	Street Address, City, State 2505 E Chestnut Ave, Suite 2b, Vineland, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6091	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(4)(iii)</p> <p>The laboratory director must ensure all proficiency testing reports received are reviewed by the appropriate staff to evaluate the laboratory's performance and to identify any problems that require corrective action.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the Biannual Assessment (BA) records and interview with the Testing Personnel (TP), the Laboratory Director failed to ensure BA was performed accurately in the calendar years 2017 and 2018. The findings include: 1) Second event 2017 and first and second event 2018 did not have the reviewing physician's diagnosis. 2) Second event 2017 was dated 2/23/19 by the sending physician. 3) The TP confirmed on 2/26/19 at 10:00 am the BA was not performed accurately.</p>