

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D1066047	(X3) Date Survey Completed 12/04/2018
Name of Provider or Supplier Titan Health Partners Llc	Street Address, City, State 1200 Us Highway 22 East, Bridgewater, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5401	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the Procedure Manual (PM) and interview with the Technical Consultant (TC), the laboratory failed to follow the procedure for Critical Values (CV) in October 2018. The finding includes: 1. The CV procedure stated "CV must be documented in the patients' medical record along with the physician or nurse practitioner who was notified" but there was not documented evidence the PM was followed for patient 436245 who had a Critical White Blood cell count. 2. The TC confirmed on 12/4/18 at 11:10 am that the CV procedure was not followed.</p>