

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  31D1066063	<b>(X3) Date Survey Completed</b>  10/08/2019
<b>Name of Provider or Supplier</b>  American Women's Services	<b>Street Address, City, State</b>  333 E Jimmie Leeds Road, Galloway Township, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2007</b>	<p><b>TESTING OF PROFICIENCY TESTING SAMPLES</b> CFR(s): 493.801(b)(1)</p> <p>The samples must be examined or tested with the laboratory's regular patient workload by personnel who routinely perform the testing in the laboratory, using the laboratory's routine methods</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the Proficiency Testing (PT) records and interview with the Office Manager (OM), the laboratory failed to ensure that Testing Personnel (TP) who performed Rhesus Factor Tests participated in the American Association of Bioanalysts PT events in the calendar years 2018 and 2019. The finding includes: 1. A review of all PT event revealed that one out of five TP performed PT events in 2018 and 2019. 2. The OM confirmed on 10/8/19 at 1:30 pm that PT events were not rotated between TP.</p>