

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D1076804	(X3) Date Survey Completed 12/13/2019
Name of Provider or Supplier Children's Specialized Hospital	Street Address, City, State 200 Somerset Street, New Brunswick, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D3031	<p>RETENTION REQUIREMENTS CFR(s): 493.1105(a)(3)</p> <p>Analytic systems records. Retain quality control and patient test records (including instrument printouts, if applicable) and records documenting all analytic systems activities specified in 493.1252 through 493.1289 for at least 2 years.</p> <p>This STANDARD is not met as evidenced by: Based on lack of Patient, and Quality Control (QC) printouts and interview with the Laboratory Director (LD), the laboratory failed to retain iStat analyzer printouts for Routine Chemistry tests from 5/28/19 the date of survey. The LD confirmed on 12/13 /19 at 10:30 am that analyzer printouts were not retained.</p>