

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D1078763	(X3) Date Survey Completed 10/10/2019
Name of Provider or Supplier Advanced Dermatology, Pc	Street Address, City, State 570 Egg Harbor Road, Sewell, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the Competency Assessment (CA) records and interview with the Office Manage (OM), the laboratory failed to perform a CA on two out of two TP in the calendar years 2018 and 2019. The OM confirmed on 10/10/19 at 10:00 am that CA was not performed on TP in 2018 and 2019.</p>
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the Biannual Assessment (BA) records and interview with the Office Manager (OM), the laboratory failed to verify the accuracy of Mohs testing twice annually in the Calendar year 2018. The OM confirmed on 10/10/19 at 10:45 am that the laboratory did not verify the accuracy of Mohs testing.</p>