

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D1096817	(X3) Date Survey Completed 02/13/2019
Name of Provider or Supplier Clark Dermatology	Street Address, City, State 1075 Central Avenue, Clark, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5401	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the Procedure Manual (PM) and interview with the Office Manager (OM), the laboratory failed to have a procedure for slide retention of Moh's slides from 1/26/17 to the date of the survey. The OM confirmed on 2/13/19 at 10:00 am that the laboratory did not have a procedure for slide retention of Moh's slides.</p>