Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D1096817	(X3) Date Survey Completed 09/13/2021
Name of Provider or Supplier Clark Dermatology	**Street Address, City, State** 1075 Central Avenue, Clark, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235 As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency. This STANDARD is not met as evidenced by: Based on surveyor review of the Competency Assessment (CA) records and interview with the Office Manager (OM), the laboratory failed to perform the CA for Testing Personnel (TP) in 2020.The finding includes: 1. One out of two TP did not have a CA performed in the calendar year 2020. 2. The OM confirmed on 9/13/21 at 12:30 pm that the CA was not performed.