

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  31D1096817	<b>(X3) Date Survey Completed</b>  12/11/2024
<b>Name of Provider or Supplier</b>  Clark Dermatology	<b>Street Address, City, State</b>  1075 Central Avenue, Clark, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor observation of Tissue Marking Dyes and Reagents, surveyor review of the Procedure Manual (PM) and interview with the Office Manager (OM) the laboratory used and failed to discard expired Tissue Marking Dye used for Histopathology testing from 6/30/24 to 12/11/24. The findings include: 1. One Green Tissue Marking Dye Lot# 154057 was observed to be expired on 7/31/24. 2. One Black Tissue Marking Dye Lot# 153190 was observed to be expired on 6/30/24. 3. Approximately 85 tests were performed and reported with the expired dyes. 4. The OM confirmed on 12/11/24 at 1:35 pm that the laboratory used and failed to discard expired dyes for Histopathology testing.</p>