

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D1104280	(X3) Date Survey Completed 07/26/2022
Name of Provider or Supplier Dermatology And Skin Surgery Center, Pa	Street Address, City, State 205 Ridgedale Avenue, Suite 4, Florham Park, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5401	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the Procedure Manual (PM), observation of Staining Station (SS) and interview with the Testing Personnel (TP), the laboratory failed to follow the PM for Hematoxylin-Eosin (HE) staining from 7/23/19 to the date of the survey. The findings include: 1. The SS in the laboratory did not correspond with the staining procedure in the PM. 2. The observation of the staining solutions and reagents were as follows: Scotts blueing, Hematoxylin, Acid alcohol, 95% alcohol, "unlabeled" staining jar, for stations one through five. 3. The PM stated Hemotoxylin, Water Rinse, Acid Alcohol, Water Rinse, Scotts Bluing solution, for stations one through five. 4. The TP confirmed on 7/26/22 at 1:30 pm that the laboratory did not follow the PM.</p>