

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D1106287	(X3) Date Survey Completed 07/14/2022
Name of Provider or Supplier Schweiger Dermatology Group	Street Address, City, State 310 Madison Avenue, Suite 206, Morristown, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5401	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the Procedure Manual (PM), observation of staining station and interview with the Office Manager (OM), the laboratory failed to have a procedure for Hematoxylin-Eosin (HE) staining from March 2022 to the date of the survey. The OM confirmed on 7/14/22 at 1:30 pm that the laboratory did not have a procedure for HE staining.</p>