

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D2000860	(X3) Date Survey Completed 08/19/2025
Name of Provider or Supplier Premier Brain And Spine Llc	Street Address, City, State 2500 Morris Avenue, Union, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5807	<p>TEST REPORT CFR(s): 493.1291(d)</p> <p>(d) Pertinent "reference intervals" or "normal" values, as determined by the laboratory performing the tests, must be available to the authorized person who ordered the tests and, if applicable, the individual responsible for using the test results.</p> <p>This STANDARD is not met as evidenced by: Based on the surveyor review of Patient Test Report (PTR), Procedure Manual (PM), and interview with the General Supervisor (GS), the laboratory failed to have accurate Reference Intervals (RI) for Urine Quantitative Toxicology tests from 2/1/25 to 8/19/25. The findings include: 1. The analyte Norbuprenorphine had a cutoff value of 10 in the PM. 2. Surveyor review of PTR revealed Norbuprenorphine had a cutoff value of 20 on five out of five PTR. 3. The GS confirmed on 8/19/25 at 10:20 am, the laboratory did not have accurate RI.</p>