

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D2004346	(X3) Date Survey Completed 01/23/2018
Name of Provider or Supplier Dermatology Center Of New Jersey,Pc, The	Street Address, City, State 745 Us Hwy 202/206, Ste 102, Bridgewater, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6103	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(13)</p> <p>The laboratory director must ensure that policies and procedures are established for monitoring individuals who conduct preanalytical, analytical, and postanalytical phases of testing to assure that they are competent and maintain their competency to process specimens, perform test procedures and report test results promptly and proficiently, and whenever necessary, identify needs for remedial training or continuing education to improve skills.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the laboratory procedures and interview with the Office Manager (OM), the Laboratory Director failed to establish a Competency Assessment (CA) procedure with the applicable elements for Mohs Technicians from 11/10/15 to the date of the survey. The GS confirmed on 1/23/18 at 2:10 pm that a CA procedure was not established.</p>