

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D2024863	(X3) Date Survey Completed 04/23/2024
Name of Provider or Supplier Dermatology And Skin Surgery Center, Pa	Street Address, City, State 3322 Route 22 West, Branchburg, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on the review of Biannual Assessment (BA) records and interview with the Testing Personnel (TP), the laboratory failed to verify the accuracy of Histopathology testing in the calendar year 2022. The TP confirmed on 4/23/24 at 1:30 pm that the laboratory did not perform BA for Histopathology in the calendar year 2022.</p>
D5805	<p>TEST REPORT CFR(s): 493.1291(c)</p> <p>The test report must indicate the following: (c)(1) For positive patient identification, either the patient's name and identification number, or a unique patient identifier and identification number. (c)(2) The name and address of the laboratory location where the test was performed. (c)(3) The test report date. (c)(4) The test performed. (c)(5) Specimen source, when appropriate. (c)(6) The test result and, if applicable, the units of measurement or interpretation, or both. (c)(7) Any information regarding the condition and disposition of specimens that do not meet the laboratory's criteria for acceptability.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the Test Reports (TR) and interview with the Testing Personnel (TP), the laboratory failed to ensure that the TR included all the required information on the date of survey. The finding includes: 1. The TR did not have the correct name and address of where the Technical Component for histopathology was</p>

performed on two out of three TR. 2. The TP confirmed on 4/23/24 at 1:30 pm that the TR failed to include all the required information.