

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  31D2024863	<b>(X3) Date Survey Completed</b>  05/13/2026
<b>Name of Provider or Supplier</b>  Dermatology And Skin Surgery Center, Pa	<b>Street Address, City, State</b>  3322 Route 22 West, Branchburg, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5401</b>	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>(a) A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor Procedure Manual (PM) and interview with the Office Manager (OM), the laboratory failed to have to follow the procedure "Cryostat and Microscope maintained" from 4/23/24 to 5/13/26. The findings include: 1. The PM stated "Cryostat and microscope maintained yearly by Micronix" 2. There was no documented evidence that Micronix performed maintenance on the microscope Serial number 1873529. 3. The OM confirmed on 5/13/26 at 12:45 PM the laboratory did not follow the procedure "Cryostat and Microscope maintained".</p>