

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D2029686	(X3) Date Survey Completed 07/14/2026
Name of Provider or Supplier Advocare Mid-Jersey Pediatrics, Pa	Street Address, City, State 2 Research Way, Monroe Twp, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5891	POSTANALYTIC SYSTEMS QUALITY ASSESSMENT CFR(s): 493.1299(a) (a) The laboratory must establish and follow written policies and procedures for an ongoing mechanism to monitor, assess and, when indicated, correct problems identified in the postanalytic systems specified in 493.1291. This STANDARD is not met as evidenced by: Based on surveyor review of the Procedure Manual (PM) and interview with the Testing Personnel (TP), the laboratory failed to establish a procedure for verifying results manually entered into the Electronic Medical Record (EMR) from 5/12/24 to 7/14/26. The finding includes. 1. The laboratory failed to have a procedure to verify manually entered test results to ensure accuracy in the postanalytic systems. 2. The TP #13 listed on CMS from 209 confirmed on 7/14/26 at 12:00 pm that the laboratory failed to have a procedure for verifying manually entered results in to the EMR.