

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D2041037	(X3) Date Survey Completed 02/25/2025
Name of Provider or Supplier Titan Health Partners Llc	Street Address, City, State One Exchange Place, Jersey City, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5801	<p>TEST REPORT CFR(s): 493.1291(a)</p> <p>(a) The laboratory must have an adequate manual or electronic system(s) in place to ensure test results and other patient-specific data are accurately and reliably sent from the point of data entry (whether interfaced or entered manually) to final report destination, in a timely manner. This includes the following: (a)(1) Results reported from calculated data. (a)(2) Results and patient-specific data electronically reported to network or interfaced systems. (a)(3) Manually transcribed or electronically transmitted results and patient-specific information reported directly or upon receipt from outside referral laboratories, satellite or point-of-care testing locations.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the Final Report (FR) Work Records (WR), Electronic Medical Records (EMR) and interview with the Laboratory Coordinator (LC) the laboratory failed to ensure test results were reported accurately for Compete Blood Count (CBC) testing into the EMR on 2/25/25. The finding includes: 1. A review of five EMR entries revealed one patient had results as follows: a) White Blood Cell (WBC) on the WR was $10.6 \times 10^3/\text{mm}^3$ but the FR had $7.7 \times 10^3/\text{mm}^3$ b) Red Blood Cell Count (RBC) on the WR was $4.49 \times 10^6/\text{mm}^6$ but the FR had $4.48 \times 10^6/\text{mm}^6$ c) Hemoglobin (Hgb) on the WR was 12.8 g/dL but the FR had 12.4 g/dL d) Hematocrit (HCT) on the WR was 38.5 % but the FR had 37.9 % e) Platelets (PLT) on the WR was $133 \times 10^3/\text{mm}^3$ but the FR had $155 \times 10^3/\text{mm}^3$ 2. The LC confirmed on 2/25/25 at 1:30 pm that the laboratory did not ensure test results were accurately recorded in the EMR.</p>