

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D2043828	(X3) Date Survey Completed 10/10/2019
Name of Provider or Supplier Nj Certified Dermatology Pc	Street Address, City, State 668-670 Broadway, Bayonne, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5601	<p>HISTOPATHOLOGY CFR(s): 493.1273(a)(f)</p> <p>(a) As specified in 493.1256(e)(3), fluorescent and immunohistochemical stains must be checked for positive and negative reactivity each time of use. For all other differential or special stains, a control slide of known reactivity must be stained with each patient slide or group of patient slides. Reactions of the control slide with each special stain must be documented. (f) The laboratory must document all control procedures performed, as specified in this section.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of Quality Control (QC) records and interview with the Testing Personnel (TP), the laboratory failed to document Hematoxylin and Eosin (H&E) control slide reactions for Mohs testing in the calendar year 2019. The findings include: 1. The laboratory did not document H&E stain QC reactions in January, February, March, April and July of 2019. 2. The laboratory read and reported 32 patients slides in the above time period. 3. The TP #1 listed on CMS form 209 confirmed on 10/10/19 at 1:30 pm that the laboratory did not document H&E QC stain reaction.</p>